

# SEA-BIRD ELECTRONICS, INC.

13431 NE 20th Street, Bellevue, Washington, 98005-2010 USA

Phone: (425) 643 - 9866 Fax (425) 643 - 9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 0039  
CALIBRATION DATE: 11-Mar-11

WEBB GLIDER PRESSURE CALIBRATION DATA  
508 psia S/N 5631

**COEFFICIENTS:**

PA0 =	1.037910e-002	PTCA0 =	5.091686e+001
PA1 =	2.445066e-002	PTCA1 =	2.182167e-001
PA2 =	1.595721e-009	PTCA2 =	-8.563787e-003
PTHA0 =	-8.438315e+001	PTCB0 =	2.511137e+001
PTHA1 =	4.736149e-002	PTCB1 =	-9.250000e-004
PTHA2 =	-3.880209e-007	PTCB2 =	0.000000e+000

**PRESSURE SPAN CALIBRATION**

PRESSURE PSIA	INST OUTPUT	THERMISTOR OUTPUT	COMPUTED PRESSURE	ERROR %FSR
14.67	651.0	2319.0	14.69	0.00
104.93	4335.6	2320.0	104.88	-0.01
204.93	8419.7	2319.0	204.91	-0.00
304.95	12501.5	2319.0	304.93	-0.00
404.94	16580.2	2320.0	404.94	-0.00
504.95	20656.4	2321.0	504.93	-0.00
404.93	16581.1	2320.0	404.96	0.01
304.93	12502.2	2320.0	304.95	0.00
204.93	8421.3	2319.0	204.95	0.00
104.94	4337.6	2319.0	104.93	-0.00
14.67	651.0	2319.0	14.69	0.00

**THERMAL CORRECTION**

TEMP ITS90	PRESS TEMP	INST OUTPUT
32.50	2520.00	669.94
29.00	2442.90	671.05
24.00	2332.80	672.28
18.50	2212.50	672.96
15.00	2135.70	673.33
4.50	1906.80	672.61
1.00	1830.00	672.21

  

TEMP (ITS90)	SPAN (mV)
-5.00	25.12
35.00	25.08

$$y = \text{thermistor output}; t = P_{TEMPA0} + P_{TEMPA1} * y + P_{TEMPA2} * y^2$$

$$x = \text{pressure output} - P_{TCA0} - P_{TCA1} * t - P_{TCA2} * t^2$$

$$n = x * P_{TCB0} / (P_{TCB0} + P_{TCB1} * t + P_{TCB2} * t^2)$$

$$\text{pressure (psia)} = P_{A0} + P_{A1} * n + P_{A2} * n^2$$

Date, Avg Delta P %FS

11-Mar-11 0.00

