

PO Box 518  
620 Applegate St.  
Philomath, OR 97370



(541) 929-5650  
Fax (541) 929-5277  
[www.wetlabs.com](http://www.wetlabs.com)

## SLC Testing Certification

Date **5/16/2017**

S/N# **FLBBCDSLC-4728**

---

### Low temperature test #1

Chill 2.5 hr at -20 °C

### High temperature test #1

 Verify operation post-testing

Heat 2.5 hr at 50 °C

### Low temperature test #2

 same protocol as #1

### High temperature test #2

 same protocol as #1

### Vacuum test

< 0.1" Hg change in 10 min.

### Pressure test

5 cycles, 0–1250 m with 10-sec. soaks

Held at 1250 m for 2 hrs. on last cycle

### Electrical isolation

Resistance between copper faceplate and grounding wire is > 1 mΩ

### Calibration verification

Verify calibration and dark counts in bb, chl, and CDOM channels

Verify 5% of single point check for chl and bb

Verify 10% of single point check for CDOM

**Signature** William Jason Rowe/Ann Gaidos-Morgan

---

**NOTES:**