



Sea-Bird Scientific
 13431 NE 20th Street
 Bellevue, WA 98005
 USA

+1 425-643-9866
 seabird@seabird.com
 www.seabird.com

SENSOR SERIAL NUMBER: 9324
 CALIBRATION DATE: 27-Feb-21

Slocum Payload CTD CONDUCTIVITY CALIBRATION DATA
 PSS 1978: C(35,15,0) = 4.2914 Siemens/meter

COEFFICIENTS:

g = -1.001298e+00 CPcor = -9.5700e-008
 h = 1.358298e-01 CTcor = 3.2500e-006
 i = -8.499083e-05 WBOTC = 5.2164e-07
 j = 2.495408e-05

BATH TEMP (° C)	BATH SAL (PSU)	BATH COND (S/m)	INSTRUMENT OUTPUT (Hz)	INSTRUMENT COND (S/m)	RESIDUAL (S/m)
22.0000	0.0000	0.00000	2715.54	0.00000	0.00000
1.0000	34.6533	2.96336	5397.30	2.96335	-0.00001
4.4999	34.6342	3.26922	5600.88	3.26925	0.00002
15.0000	34.5944	4.24726	6206.33	4.24720	-0.00006
18.5000	34.5863	4.59114	6405.42	4.59117	0.00003
24.0000	34.5779	5.14707	6714.45	5.14709	0.00003
29.0001	34.5740	5.66709	6990.84	5.66708	-0.00002
32.4999	34.5716	6.03812	7181.27	6.03791	-0.00021

$$f = \text{Instrument Output(Hz)} * \text{sqrt}(1.0 + \text{WBOTC} * t) / 1000.0$$

t = temperature (°C); p = pressure (decibars); δ = CTcor; ϵ = CPcor;

$$\text{Conductivity (S/m)} = (g + h * f^2 + i * f^3 + j * f^4) / (1 + \delta * t + \epsilon * p)$$

Residual (Siemens/meter) = instrument conductivity - bath conductivity

